



Number **TC7349** revision 0  
Project number 710915  
Page 1 of 4

Issued by NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

Notified Body Number 0122

In accordance with Paragraph 8.1 of the European Standard on Metrological aspects of non-automatic weighing instruments EN 45501:1992/AC:1993 and by application of the OIML International Recommendation R 60 (Edition 2000). The applied error fraction  $p_i$ , meant in the paragraph 3.5.4. of the standard is 0.7.

Applicant Keli Electric Manufacturing (Ningbo) Co., Ltd.  
No. 199 Changxing Road,  
Jiangbei District, Ningbo,  
P.R. China

In respect of A **shear beam load cell**, with strain gauges, tested as a part of a weighing instrument.  
Manufacturer : Keli Electric Manufacturing (Ningbo) Co., Ltd.  
Type : SB-S .....

#### Characteristics

Maximum capacity ( $E_{max}$ )	1 t up to and including 5 t
Accuracy class	C
Maximum number of load cell verification intervals ( $n_{max}$ )	3000
Ratio of minimum LC verification interval $Y = E_{max} / V_{min}$	8000
Ratio of minimum dead load output return $Z = E_{max} / (2 * DR)$	6000

In the description number TC7349 revision 0 further characteristics are described.



Nederlands Meetinstituut

# Test certificate

Number **TC7349** revision 0  
Project number 710915  
Page 2 of 4

Description and documentation The load cell is described in the description number TC7349 revision 0 and documented in the documentation folder TC7349-1, appertaining to this test certificate.

Remarks Summary of the test involved: see Appendix number TC7349 revision 0

Dordrecht, 28 January 2008  
NMI Certin B.V.



Ing. C. Oosterman  
Manager Product Certification

## 1 General information about the load cell

All properties of the load cell, whether mentioned or not, may not be in conflict with the standard mentioned in the test certificate.

### 1.1 Essential parts

Description	Drawing number	Rev.	Remarks
SB-S Load cell Specifications	1710915	0	Mechanical
Electrical diagram of the SB-S load cell	2710915	0	Electrical 4 wire-system
Electrical diagram of the SB-S load cell	3710915	0	Electrical 6 wire-system

Cable:

- The load cell is provided with a 4-wire system.  
The cable length has to be as indicated on the drawing number 4710915;  
The cable length shall not be modified.
- The load cell is provided with a 6-wire system (=“Remote-sensing”).  
The cable length is not limited.
- The cable should be a shielded cable, the shield is not connected to the load cell.

### 1.2 Essential characteristics

Minimum dead load	: 0 kg
Safe overload	: 120 % of $E_{max}$
Rated Output	: 2 mV/V $\pm$ 0.002 mV/V
Input impedance	: 400 $\Omega$ $\pm$ 10 $\Omega$
Output impedance	: 350 $\Omega$ $\pm$ 2 $\Omega$
Recommended excitation	: 10 / 12 V DC
Excitation maximum	: 15 V DC
Transducer material	: Stainless Steel
Atmospheric protection	: Metallic hermetical welded

### 1.3 Essential shapes

The load cell is built according to the drawing:

- SB-S Load cell Specifications, drawing number 1710915
- Electrical diagram of the load cell, drawing numbers 2710915 and 3710915.

The data plate is secured against removal by sealing or will be destroyed when removed. The data plate mentions at least the information and markings as described in the OIML R60 document. In the countries where it is mandatory the load cell should bear this test certificate number: TC7349.

Securing:

The connecting cable of the load cell or the junction box is provided with possibility to seal.

Number **TC7349** revision 0  
Project number 710915  
Page 4 of 4

Tests performed for this test certificate:

Test	Institute	type, version, remarks
Temperature test and repeatability (20, 40, -10 and 20 °C)	NMi Certin B.V.	SB-S 1t-C3 4-wire system
Temperature effect on minimum dead load output (20, 40, -10 and 20 °C)	NMi Certin B.V.	SB-S 1t-C3 4-wire system
Creep (20, 40 and -10 °C)	NMi Certin B.V.	SB-S 1t-C3 4-wire system
Minimum dead load output return (20, 40 and -10 °C)	NMi Certin B.V.	SB-S 1t-C3 4-wire system
Barometric pressure effects at room temperature	NMi Certin B.V.	SB-S 1t-C3 4-wire system
Damp heat, cyclic: marked CH (or not marked)	NMi Certin B.V.	SB-S 1t-C3 4-wire system