



Number **TC6992** revision 0
Project number 604389
Page 1 of 4

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

Notified Body Number 0122

In accordance with Paragraph 8.1 of the European Standard on Metrological aspects of non-automatic weighing instruments EN 45501:1992/AC:1993 and by application of the OIML International Recommendation R 60 (Edition 2000). The applied error fraction p_i , meant in the paragraph 3.5.4. of the standard is 0.7.

Applicant Keli Electric Manufacturing (Ningbo) Co., Ltd.
No. 199 Changxing Road,
Jiangbei District, Ningbo City
P.R. of China

In respect of **A double ended shear beam load cell**, with strain gauges, tested as a part of a weighing instrument.
Manufacturer : Keli Electric Manufacturing (Ningbo) Co., Ltd.
Type : QS series

Characteristics

Maximum capacity (E_{max})	10 000 kg up to and including 50 000 kg
Accuracy class	C
Maximum number of load cell verification intervals (n_{max})	3000
Ratio of minimum LC verification interval $Y = E_{max} / V_{min}$	10 000
Ratio of minimum dead load output return $Z = E_{max} / (2 * DR)$	3000

In the description number TC6992 revision 0 further characteristics are described.



Nederlands Meetinstituut

Test certificate

Number **TC6992** revision 0
Project number 604389
Page 2 of 4

Description and documentation The load cell is described in the description number TC6992 revision 0 and documented in the documentation folder TC6992-1, appertaining to this test certificate.

Remarks Summary of the test involved: see Appendix number TC6992 revision 0

Dordrecht, 4 July 2006
NMI Certin B.V.


Ing. C. Oosterman
Manager Product Certification

1 General information about the load cell

All properties of the load cell, whether mentioned or not, may not be in conflict with the standard mentioned in the test certificate.

1.1 Essential parts

Description	Drawing number	Rev.	Remarks
QS load cell, 10tf	0001	0	Mechanical
Electrical diagram of the QS load cell	0002	0	Electrical

Cable:

- The load cell is provided with a 4-wire system.
The cable length for the E_{max} of:
 - 5 t is 5.2 m;
 - 10 t is 8 m;
 - 15 t is 10 m;
 - 20, 25 and 30 t is 12 m;
 - 40 and 50 t is 16 m.
 The cable length shall not be modified.
- The cable should be a shielded cable, the shield is not connected to the load cell.

1.2 Essential characteristics

Minimum dead load	: 0 kg
Safe overload	: 120 % of E_{max}
Rated Output	: 2 mV/V \pm 0.002 mV/V
Input impedance	: 750 Ω \pm 10 Ω
Output impedance	: 703 Ω \pm 2 Ω
Recommended excitation	: 10 / 12 V DC
Excitation maximum	: 15 V DC
Transducer material	: Alloy steel
Atmospheric protection	: IP68

1.3 Essential shapes

The load cell is built according to drawing:

- QS Load cell, 10 tf, drawing number 0001.

The data plate is secured against removal by sealing or will be destroyed when removed. The data plate mentions at least the information and markings as described in the OIML R60 document. In the countries where it is mandatory the load cell should bear this test certificate number: TC6992.

Securing:

The connecting cable of the load cell or the junction box is provided with possibility to seal.

Number **TC6992** revision 0
 Project number 604389
 Page 4 of 4

Tests performed for this test certificate:

Test	Institute	type, version, remarks
Temperature test and repeatability (20, 40, -10 and 20 °C)	NMi Certin B.V.	QS-A 10 000 kg C3
Temperature effect on minimum dead load output (20, 40, -10 and 20 °C)	NMi Certin B.V.	QS-A 10 000 kg C3
Creep (20, 40 and -10 °C)	NMi Certin B.V.	QS-A 10 000 kg C3
Minimum dead load output return (20, 40 and -10 °C)	NMi Certin B.V.	QS-A 10 000 kg C3
Barometric pressure effects at room temperature	NMi Certin B.V.	QS-A 10 000 kg C3
Damp heat, cyclic: marked CH (or not marked)	NMi Certin B.V.	QS-A 10 000 kg C3